# QUANTOFIX® Cyanide

#### Pack content:

1 aluminium container with 100 test strips

1 bottle Cyanide-1 1 measuring vessel with 5 mL mark

1 bottle Cyanide-2 1 small measuring spoon

### Measuring range:

1-30 mg/L CN

## Color gradation:

 $0 \cdot 1 \cdot 3 \cdot 10 \cdot 30 \text{ mg/L CN}^{-1}$ 

## Safety precautions:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

#### General indications:

Remove only as many test strips as are required. Close the container immediately after removing a strip. Do not touch the test field.

#### Additional indications:

To obtain correct results with this test, the pH value of the sample must be < 10. Therefore, the pH value of the sample must be first measured (for example with pH-Fix 0-14, REF 92110). If the solution has a pH value > 10, add some drops of diluted sulfuric acid to lower it to 6-7.

#### Instructions for use:

- Rinse the measuring vessel with the test solution and fill it to the 5 mL mark.
- 2. Add 1 spoon Cyanide-1 (buffer solution) to the sample.
- 3. Add 5 drops Cyanide-2 to the sample.
- 4. Shake carefully.
- 5. Immediately dip the test strip into the sample.
- 6. Wait 45 s.
- 7. Shake off excess liquid.
- 8. Compare within 10 s with the color scale.

#### Interferences:

This test detects free cyanide and cyanide complexes which are decomposed by chlorine. Thiocyanates (SCNT) above 1 mg/L lead to a similar coloration. lodide and bromide in concentrations of 5 mg/L and above lead to a fading of the color reaction, and in major concentrations may inhibit the color reaction. Sulfide ions (S $^2$ T) up to 20 mg/L do not interfere. Higher concentrations also decrease or inhibit the color reaction. If the pH value of the sample is > 10, then the capacity of the buffer solution will not be sufficient to lower the pH to the optimal value of 6–7. In this case, prior to the test, the pH value of the sample should be adjusted to 6–7 by drop wise addition of diluted sulfuric acid.

# Disposal:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

#### Storage:

Avoid exposing the strips to sunlight and moisture. Keep container cool and dry (storage temperature 4–30 °C).

If correctly stored, the test kit may be used until the use-by-date printed on the packaging.

CTL SCIENTIFIC SUPPLY CORP. 1016-3 Grand Boulevard, Deer Park, NY 11729

Tel: 631-242-4249

Web: www.ctlscientific.com

Manufacturer: Macherey-Nagel GMbH & CO. KG Rev: 2025-03















