

QUANTOFIX® Peroxide 1000

Pack content:

1 aluminum container with 100 test strips

General indications:

Remove only as many test strips as are required. Close the container immediately after removing a strip. Do not touch the test fields.

Measuring range:

Visually
50–1000 mg/L H₂O₂

Reflectometrically
50–1000 mg/L H₂O₂

Color gradation:

0 · 50 · 150 · 300 · 500 · 800 · 1000 mg/L H₂O₂

Additional indications:

If hydrogen peroxide is present, the test field at the end of the strip turns yellow to brown. If the warning field turns blue during the reaction time of 45 seconds, the hydrogen peroxide concentration is greater than 2000 mg/L, although the peroxide test field may indicate a value within the measuring range. In this case, we recommend carrying out a stepwise dilution of the sample until the warning field no longer reacts and a hydrogen peroxide concentration below 1000 mg/L is measured. The test fields may change after the reaction time has elapsed. This has no relevance for the reading.

Instructions for use:

1. Dip the test strip with both fields into the test solution for 1 second.
2. Shake off excess liquid.
3. Wait 45 seconds.
4. Compare with the color scale.

Interferences:

If the sample solution has a pH value of 2–9, the reaction will take place without interferences. Strongly acidic solutions must be buffered with sodium acetate, and alkaline solutions with citric acid to a pH of 5–7. Only the presence of strong oxidants interferes with the determination.

The following ions interfere with the determination only in higher concentrations:

- > 1000 mg/L: formaldehyde (HCHO)
- > 500 mg/L: Ca²⁺, NO₃⁻
- > 100 mg/L: ascorbate, free chlorine, combined chlorine, SO₃²⁻
- > 70°d: total hardness

If free and combined chlorine is present, the warning field turns blue.





Disposal:

Used test strips can be placed in the normal household waste.

Storage:

Avoid exposing the strips to sunlight and moisture. Keep container cool and dry (storage temperature not above +30 °C).

If correctly stored, the test strips may be used until the use-by-date printed on the packaging.



Product data and ordering information

Type	Semi-quantitative test strips
Presentation	Tube of 100 test strips
Color reaction	white to brown
Gradation	0 · 50 · 150 · 300 · 500 · 800 · 1000 mg/l (ppm) H ₂ O ₂ + Warning field
REF	91333