

REF 91820

Test 1-20

03.22

NANOCOLOR® Chloride**Method:**

Photometric determination with mercury(II) thiocyanate and iron(III) nitrate

Cuvette rectangular:	50 mm	10 mm
Range (mg/L Cl ⁻):	0.2–20.0	1–125
Wavelength (HW = 5–12 nm):	470 nm	
Reaction time:	1 min (60 s)	
Reaction temperature:	20–25 °C	

Contents of reagent set:

2 x 100 mL Chloride R1

2 x 100 mL Chloride R2

Hazard warning:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

Preliminary tests:

If the order of magnitude of the concentration in a sample is not known, a preliminary test with QUANTOFIX® Chloride (500–3000 mg/L Cl⁻, REF 91321) or with VISOCOLOR® HE Chloride CL 500 (REF 915004) rapidly gives this information. From the order of magnitude the required dilution can be calculated and prepared directly.

Interferences:

Thiocyanate, sulfide, thiosulfate, bromide and iodide all interfere since they react in the same way as chloride. A fluoride concentration in excess of 20 mg/L interferes in the chloride determination, and the concentrations read off are lower than those actually present in the test sample.

The method cannot be applied for the analysis of sea water.

Procedure:

Requisite accessories: volumetric flasks 25 mL, piston pipette with tips

Pour into two separate volumetric flasks 25 mL:

Test sample	Blank value
20 mL test sample (<i>the pH value of the sample must be between pH 4 and 13</i>)	20 mL distilled water
2 mL R1, mix	2 mL R1, mix
2 mL R2, mix	2 mL R2, mix

Fill up test sample and the blank value to 25 mL mark with distilled water and mix again. After 1 min pour into cuvettes and measure.

Measurement:

For MACHEREY-NAGEL photometers see manual, test 1-20.

Measurement when samples are colored or turbid:

For all NANOCOLOR® photometers see manual, use key for correction value.

Photometers of other manufacturers:

Verify calibration curve for each type of instrument by measuring standard solutions.

Analytical quality control:

NANOCONTROL Multistandard Drinking Water (REF 925018)

Decreasing volume of analytical preparation:

In order to increase the number of determinations, you can work with volumetric flasks of 10 mL: 8 mL test sample + 0.8 mL R1 + 0.8 mL R2, semi-micro cuvette (REF 91950).

Disposal:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

CTL SCIENTIFIC SUPPLY CORP. 1016-3 Grand Boulevard, Deer Park, NY 11729

Tel: 631-242-4249

Web: www.ctlscientific.com

Manufacturer: Macherrey-Nagel GmbH & Co. KG

REV: 2025-09