

REF 91830

Test 1-30 10.22

NANOCOLOR® Cyanide

Method:

Photometric determination with barbituric acid / pyridine

Cuvette:	50 mm	10 mm
Range (mg/L CN ⁻):	0.001 – 0.100	0.01 – 0.50
Wavelength (HW = 5 – 12 nm):	585 nm	
Reaction time:	5 min (300 s)	
Reaction temperature:	20 – 25 °C	

Contents of reagent set:

7 g Cyanide R1 1 small measuring spoon 70 mm
 12 g Cyanide R2 1 Large measuring spoon 85 mm
 2 × 100 mL Cyanide R3

Hazard warning:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

Interferences:

Thiocyanate interferes by reacting the same way as cyanide (determination with test 0-90 NANOCOLOR® Thiocyanate 50).

Only free cyanide and cyanide complexes which can be decomposed by chlorine are determined.

When interfering substances, such as heavy metal complexes, thiocyanate, sulfide, dyes or aromatic amines are present, cyanide must be separated by distillation before determination (see „Note“).

The method can be applied also for the analysis of sea water.

Note:

For the determination of readily liberated cyanide and total cyanide please contact MACHEREY-NAGEL for special working instructions.

Procedure:

Requisite accessories: volumetric flasks 25 mL, piston pipette with tips

Pour into two separate volumetric flasks:

Test sample	Blank value
20 mL test sample (the pH value of the sample must be between pH 6 and 8)	20 mL test sample (the pH value of the sample must be between pH 6 and 8)
1 level small spoon R1, dissolve wait 1 min	–
1 level large spoon R2, mix thoroughly (some grains remain undissolved)	–
2 mL R3, mix	–

Test sample (< 0.02 mg/L CN ⁻)	Blank value
20 mL test sample (the pH value of the sample must be between pH 6 and 8)	20 mL distilled water
1 level small spoon R1, dissolve wait 1 min	1 level small spoon R1, dissolve wait 1 min
1 level large spoon R2, mix thoroughly (some grains remain undissolved)	1 level large spoon R2, mix thoroughly (some grains remain undissolved)
2 mL R3, mix	2 mL R3, mix

Fill up sample and blank value to 25 mL mark with distilled water and mix again. After 5 min pour into cuvettes and measure.

Measurement:

For MACHEREY-NAGEL photometers see manual, test 1-30.

Photometers of other manufacturers:

Verify factor for each type of instrument by measuring standard solutions.

Decreasing volume of analytical preparation:

In order to increase the number of determinations, you can work with volumetric flasks of 10 mL: 8 mL test sample +½ level spoon R1 +½ level spoon R2 +0.8 mL R3, semi micro cuvette (REF 91950).

Disposal:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

CTL SCIENTIFIC SUPPLY CORP. 1016-3 Grand Boulevard, Deer Park, NY 11729

Tel: 631-242-4249

Web: www.ctlscientific.com

Manufacturer: Macherey-Nagel GmbH & Co. KG

REV: 2025-09