

REF 91878

Test 1-78

01.22

NANOCOLOR® ortho-Phosphate

Method:

Photometric determination of the yellow phosphate-molybdate-vanadate complex in acidic solution

Cuvette:	50 mm	10 mm
Range (mg/L PO_4^{3-}):	0.5–20.0	2–50
Range (mg/L $\text{PO}_4\text{-P}$):	0.2–6.6	1–17
Wavelength (HW = 5–12 nm):	436 nm	
Reaction time:	10 min (600 s)	
Reaction temperature:	20–25 °C	

Contents of reagent set:

2 x 100 mL o-Phosphate R1

2 x 100 mL o-Phosphate R2

Hazard warning:

Information regarding safety can be found on the box' label and in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

Interferences:

For the determination of total phosphorus use tests 0-55, 0-76, 0-79, 0-80 or 0-81.

The following quantities of ions will not interfere: $\leq 10 \text{ mg/L Fe}$, $\leq 1000 \text{ mg/L Si}$.

The method can be applied also for the analysis of sea water.

Note:

Please contact MACHEREY-NAGEL for special working instructions concerning a simplified procedure in a beaker (without filling up) and evaluation in 50 mm cuvette.

Procedure:

Requisite accessories: volumetric flasks 25 mL, piston pipette with tips

Pour into two separate volumetric flasks 25 mL:

Test sample	Blank value
20 mL test sample (<i>the pH value of the sample must be between pH 1 and 13</i>)	20 mL distilled water
1 mL R1, mix	1 mL R1, mix
1 mL R2, mix	1 mL R2, mix

Fill up sample and blank value to 25 mL mark with distilled water and mix again. After 10 min pour into cuvettes and measure.

Measurement:

For NANOCOLOR® photometers see manual, test 1-78.

Measurement when samples are colored or turbid:

For all NANOCOLOR® photometers see manual, use key for correction value.

Photometers of other manufacturers:

Verify factor for each type of instrument by measuring standard solutions.

Decreasing volume of analytical preparation:

In order to increase the number of determinations, you can work with volumetric flasks of 10 mL: 8 mL test sample + 0.4 mL R1 + 0.4 mL R2, semi-micro cuvette (REF 91950).

Disposal:

Information regarding disposal can be found in the safety data sheet. You can download the SDS from www.mn-net.com/SDS.

CTL SCIENTIFIC SUPPLY CORP. 1016-3 Grand Boulevard, Deer Park, NY 11729

Tel: 631-242-4249

Web: www.ctlscientific.com

Manufacturer: Macherey-Nagel GmbH & Co. KG

REV: 2025-09