

REF 985091

Test 0-91

10.24

NANOCOLOR® Thiocyanate 50

Method:

Photometric determination as iron(III) thiocyanate

Range :	0.5 – 50.0 mg/L SCN ⁻
Wavelength (HW = 5 – 12 nm):	470 nm
Reaction time:	0 min
Reaction temperature:	20 – 25 °C

Contents of reagent set:

20 test tubes Thiocyanate 50

Hazard warning:

Test tubes contain Hydrochloric acid 10 – 25 %.

For further information ask for a safety data sheet.

Interferences:

Nitrite, fluoride, anions of organic acids, phosphate, arsenate and borate interfere due to formation of complexes.

The method can also be applied for the analysis of sea water after dilution (1 + 1).

Procedure:

Requisite accessories: piston pipette with tips

Open test tube, add

4.0 mL test sample (the pH value of the sample must be between pH 7 and 13), close and mix.

Clean outside of test tube and measure immediately.

Note:

This method can also be used for detecting a thiocyanate interference in the cyanide test, mg/L SCN⁻ roughly corresponds to x mg/L CN⁻:

mg/L SCN ⁻	mg/L CN ⁻ (Test 0-31)
0.1	0.05
0.5	0.13
1.0	0.23
1.5	0.34
2.0	0.43

Measurement:

For MACHEREY-NAGEL photometers see manual, test 0-91.

Measurement when samples are colored or turbid:

For all NANOCOLOR® photometers see manual, use key for correction value.

Photometers of other manufacturers:

For other photometers check whether measurement of round glass tubes is possible. Verify factor for each type of instrument by measuring standard solutions.

CTL SCIENTIFIC SUPPLY CORP. 1016-3 Grand Boulevard, Deer Park, NY 11729

Tel: 631-242-4249

Web: www.ctlscientific.com

Manufacturer: Macherey-Nagel GmbH & CO. KG

Rev: 2025-03