

Ammonium 3

Test kit for performing colorimetric tests
on ammonium ions in surface water and sewage

Method:

Monochloramine is derived from ammonium ions as a result of the effect chlorine in the alkaline range. Combined with thymol, this forms a blue indo phenol dye.

Measurement range:

0.2–3 mg/L NH₄⁺

Contents of test kit (*refill pack):

sufficient for 50 tests

- 30 mL NH₄-1*
- 2.5 g NH₄-2*
- 6 mL NH₄-3*
- 1 measuring spoon 70 mm*
- 2 screw-plug measuring glasses
- 1 slide comparator
- 1 color chart
- 1 plastic syringe 5 mL
- 1 instructions for use*

Hazard warning:

NH₄-1 contains sodium hydroxide solution 5–20%, NH₄-3 contains ethanol 35–55% and thymol 5–10%.

H314 Causes severe skin burns and eye damage.

P260, P280, P301+330+331, P303+361+353, P304+340, P305+351+338,

P501 Do not breathe vapors. Wear protective gloves/eye protection. IF

SWALLOWED: Rinse mouth. Do NOT induce vomiting. IF ON SKIN (or hair):

Take off immediately all contaminated clothing. Rinse skin with water/shower.

IF INHALED: Remove person to fresh air and keep comfortable for breathing.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact

lenses, if present and easy to do. Continue rinsing. Dispose of contents/con-

tainer to regulated waste treatment. For further information ask for a safety

data sheet.

Instructions for use:

also refer to the pictogram on the back of the color chart

1. Pour a **5 mL** water sample into each of the measuring glasses using the plastic syringe. Place a measuring glass on position A in the comparator.

Only add the reagent to measuring glass B.

2. Add **10 drops of NH₄-1**. Seal the glass and mix.
3. Add **1 level measuring spoonful of NH₄-2**, seal the glass and shake the mixture until the powder has dissolved. Wait for **5 min**.
4. Add **4 drops of NH₄-3**. Seal the glass and mix.
5. Open the glass after **7 min** and place it on position B in the comparator.
6. Slide the comparator until the colors match in the inspection hole on top. Check the measurement reading in the recess on the comparator reed. Mid-values can be estimated.
7. After use, rinse out both measuring glasses thoroughly and seal them.

The reagents can be used for the **photometric evaluation** with photometer PF-12.

This technique can also be used for analyzing sea water after dilution (1+9).

Disposing of the samples:

The used analysis specimens can be flushed down the drain with tap water and channelled off to the local sewage treatment works.

Interferences:

Primary amines react in the same way as ammonium ions and produce higher results.

Depending on their concentration, substances which draw on the chlorine may reduce the measurement reading or suppress the reaction totally.

The temperature of the water sample should be between 18 and 30 °C. Especially low temperatures decrease the reaction rate considerably (low results).

Conversion table:

mg/L NH ₄ ⁺	mg/L NH ₄ -N (ammonia nitrogen)
0.2	0.16
0.3	0.23
0.5	0.39
0.7	0.54
1	0.78
2	1.6
3	2.3

Storage:

Store the test kit in a cool (< 25 °C) and dry place.

Product data and ordering information

REF	931 008 (931 208)
Type	colorimetric test kit (refill pack)
Range	0 · 0.2 · 0.3 · 0.5 · 0.7 · 1 · 2 · 3 mg/l (ppm) NH ₄ ⁺
Sufficient for	50 determinations
Shelf life	at least 1.5 years
Sea water suitability	yes, after dilution (1+9)
Detectable with PF-12	yes